H.B

·	Application No.	Applicant(s)
Notice of Allowability	10/716,148	MATSUMOTO, KAZUHIRO
	Examiner	Art Unit
	Irakli Kiknadze	2882
The MAILING DATE of this communication appe All claims being allowable, PROSECUTION ON THE MERITS IS a herewith (or previously mailed), a Notice of Allowance (PTOL-85) NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RI of the Office or upon petition by the applicant. See 37 CFR 1.313 1. ☐ This communication is responsive to 2. ☐ The allowed claim(s) is/are 1-7. 3. ☐ The drawings filed on 17 November 2003 are accepted by 4. ☐ Acknowledgment is made of a claim for foreign priority un a) ☐ All b) ☐ Some* c) ☐ None of the: 1. ☐ Certified copies of the priority documents have 2. ☐ Certified copies of the priority documents have	ears on the cover sheet will (OR REMAINS) CLOSED if or other appropriate comming GHTS. This application is and MPEP 1308. the Examiner. Index 35 U.S.C. § 119(a)-(d) been received.	ith the correspondence address n this application. If not included unication will be mailed in due course. THIS subject to withdrawal from issue at the initiative or (f).
3. Copies of the certified copies of the priority doc International Bureau (PCT Rule 17.2(a)). * Certified copies not received: Applicant has THREE MONTHS FROM THE "MAILING DATE" of noted below. Failure to timely comply will result in ABANDONM THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.	of this communication to file	
 A SUBSTITUTE OATH OR DECLARATION must be submi INFORMAL PATENT APPLICATION (PTO-152) which give 		
6. ☐ CORRECTED DRAWINGS (as "replacement sheets") must (a) ☐ including changes required by the Notice of Draftspers. 1) ☐ hereto or 2) ☐ to Paper No./Mail Date (b) ☐ including changes required by the attached Examiner's Paper No./Mail Date Identifying indicia such as the application number (see 37 CFR 1. each sheet. Replacement sheet(s) should be labeled as such in the	on's Patent Drawing Review Amendment / Comment o	r in the Office action of he drawings in the front (not the back) of
 DEPOSIT OF and/or INFORMATION about the depose attached Examiner's comment regarding REQUIREMENT F 		
 Attachment(s) 1. Notice of References Cited (PTO-892) 2. Notice of Draftperson's Patent Drawing Review (PTO-948) 3. Information Disclosure Statements (PTO-1449 or PTO/SB/06 Paper No./Mail Date 11/22/2004 4. Examiner's Comment Regarding Requirement for Deposit of Biological Material 	6. ☐ Interview S Paper No 8), 7. ☐ Examiner's	nformal Patent Application (PTO-152) ummary (PTO-413), /Mail Date Amendment/Comment Statement of Reasons for Allowance
· ·		

DETAILED ACTION

Allowable Subject Matter

- 1. Claims 1-7 are allowed.
- 2. The following is an examiner's statement of reasons for allowance:

Claims 1-7 are allowed because prior art fails to teach or make obvious an X-ray digital tomographic image taking apparatus comprising: a control unit for controlling the emissions of an X-ray generation unit and the drive of solid state image sensing unit, wherein the control unit divides the course along which the X-ray generation unit or solid state image sensing unit is moving into a plurality of sections and controls the X-ray generation unit so that the quality of the x-rays differs between neighboring sections as claimed.

Any comments considered necessary by applicant must be submitted no later than the payment of the issue fee and, to avoid processing delays, should preferably accompany the issue fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for Allowance."

Conclusion

3. The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Bisek et al. (Re. 36,162); Carbal et al. (US Patent 6,102,567); Wilson et al. (US Patent 6,496, 557 B2) and Stein et al. (US Patent 5,715,820) teach dual-energy topographic image taking methods and apparatus wherein the X-ray

Application/Control Number: 10/716,148

Art Unit: 2882

generation units or solid state image sensing units are moving into a plurality of sections producing desirable images with dual-energy subtraction without changing the X-ray irradiation conditions such as tube current, tube voltage and X-ray irradiation time between neighboring sections.

Page 3

4. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Irakli Kiknadze whose telephone number is 571-272-2493. The examiner can normally be reached on 9:00- 5:30.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Ed Glick can be reached on 571-272-2490. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

Irakli Kiknadze March 18, 2005

IK